



File D:\DATA\A02\MONITOR\BKM-120-05-34972-LCMSA002.D

Injection Date : 5 May 16 1:22 pm +0800 Tgt Mass (EZ) :
 Sample Name : BKM-120 05 Location : P1-C-07
 Acq. Operator : A02-MONITOR Inj : 1
 Spec. Reported : MS Integration Inj Volume : 2 ul
 Acq. Method : D:\METHODS\1-POS-3MIN-MON.M
 Analysis Method : D:\METHODS\1-POS-3MIN-MON.M
 Sample Info : Easy-Access Method: '1-POS-3MIN-MON'
 Method Info : Mobile Phase: A: water (0.01%TFA) B:ACN(0.01%TFA)
 Gradient: 5% to 95%B within 1.6 min
 Flow Rate :2.0ml/min
 Column :SunFire C18, 4.6*50mm, 3.5um
 Oven Temperature : 50C



Integration Results for ADC1 A, ADC1 CHANNEL A


RetTim	Width	Area	Height	Area%
1.11	0.02	5.89	3.09	100.00

Integration Results for DAD1 A, Sig=214.4 Ref=off

RetTim	Width	Area	Height	Area%
1.09	0.02	254.79	188.95	100.00



Integration Results for DAD1 B, Sig=254.4 Ref=off

RetTim	Width	Area	Height	Area%
1.09	0.02	645.41	475.67	100.00



Ret. Time: 1.09

ES-API Positive



CHEMIETEK