

CHEMIETEK HPLC Report

Acq. Operator : CTK
Acq. Instrument : HPLC1

Location : P2-B-03


Inj Volume : 2.500 µl

Method : D:\METHODS\1-HPLC-15MIN-2.M



(modified after loading)
Method Info : Mobile Phase: A: water(0.05%TFA) B: ACN(0.05%TFA)
Gradient: 0min 5%B,3min 5%B,10min 95%B,15min 95%B
Flow Rate: 1.2 ml/min
Column:
Oven Temperature: 40 C

Sample Info : AZD0156 CHEMIETEK Lot 01




Area Percent Report

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 B, Sig=254,4 Ref=off

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	6.920	BB	0.0689	2222.89014	472.27393	100.0000




CHEMIETEK HPLC Report

Totals : 2222.89014 472.27393

Signal 2: DAD1 A, Sig=214,4 Ref=off

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	6.920	BB	0.0711	1383.94189	282.60989	100.0000

Totals : 1383.94189 282.60989



*** End of Report ***

www.chemietek.com