



Injection Info Default
 Injection Volume 5 uL
 Co-Solvent
 Column
 Sample AZD-4573 CHEMIETEK Lot 01
 Sample Well P1: 1E
 Column Temperature 40
 CO2 Flow Rate 2.8
 Co-Solvent Flow Rate 1.2
 Co-Solvent % 30
 Total Flow 4
 Front Pressure 150
 Back Pressure 121
 Pressure Drop 29
 PDA Start Wavelength 214
 PDA Stop Wavelength 359



Peak Number	RT (min)	Area %	Area	Height
1	2.65	0.2453	24.2409	3.5527
2	3.01	99.7547	9855.9496	670.9587